

ತುಮಕೂರು



ವಿಶ್ವವಿದ್ಯಾನಿಲಯ

ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ಕಾರ್ಯಾಲಯ, ಜಿ.ಎಚ್.ರಸ್ತೆ, ತುಮಕೂರು-572 103

ದೂರವಾಣಿ : 0818-2254548, 0818-2255588, 0818-2255597 ಫ್ಯಾಕ್ಸ್ : 0818-2270719

ಸಂಖ್ಯೆ: ತು.ಬಿ.ಸಾ.ಬಿ.ಇ.ದ.ತಾ.2019-20/2046

ದಿನಾಂಕ: 23.08.2019

ಮರು ದರಪಟ್ಟಿ ಆಹ್ವಾನ

16

ವಿಷಯ: ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ವಿಜ್ಞಾನ ಕಾಲೇಜಿನ ಸ್ನಾತಕೋತ್ತರ ಭೌತಶಾಸ್ತ್ರ ಅಧ್ಯಯನ ಮತ್ತು ಸಂಶೋಧನ ವಿಭಾಗದ ಪ್ರಯೋಗಾಲಯಕ್ಕೆ ಅಗತ್ಯವಿರುವ ಸಲಕರಣೆ ಮತ್ತು ಬಿಡಿಭಾಗಗಳನ್ನು ಸರಬರಾಜು ಮಾಡಲು ದರಪಟ್ಟಿ ಆಹ್ವಾನಿಸಿರುವ ಕುರಿತು.

ಉಲ್ಲೇಖ: ಮಾನ್ಯ ಕುಲಪತಿಗಳ ಅನುಮೋದನೆ ದಿನಾಂಕ:13.08.2019 (ಕಂಡಿಕೆ-99)

* * *

ಮೇಲ್ಕಂಡ ವಿಷಯಕ್ಕೆ ಸಂಬಂಧಿಸಿದಂತೆ, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ವಿಜ್ಞಾನ ಕಾಲೇಜಿನ ಸ್ನಾತಕೋತ್ತರ ಭೌತಶಾಸ್ತ್ರ ಅಧ್ಯಯನ ಮತ್ತು ಸಂಶೋಧನ ವಿಭಾಗದ ಪ್ರಯೋಗಾಲಯಕ್ಕೆ ಅಗತ್ಯವಿರುವ ಅನುಬಂಧ-01ರಲ್ಲಿರುವ ಸಲಕರಣೆ ಮತ್ತು ಬಿಡಿಭಾಗಗಳನ್ನು ಸರಬರಾಜು ಮಾಡುವುದಕ್ಕಾಗಿ ದರಪಟ್ಟಿಗಳನ್ನು ಈ ಕೆಳಕಂಡ ಷರತ್ತುಗಳೊಳಪಟ್ಟು ಆಹ್ವಾನಿಸಲಾಗಿದೆ.

ಷರತ್ತುಗಳು:

1. ಲಕೋಟೆಯ ಮೇಲೆ ವಿಶ್ವವಿದ್ಯಾನಿಲಯದ "ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ವಿಜ್ಞಾನ ಕಾಲೇಜಿನ ಸ್ನಾತಕೋತ್ತರ ಭೌತಶಾಸ್ತ್ರ ಅಧ್ಯಯನ ಮತ್ತು ಸಂಶೋಧನ ವಿಭಾಗದ ಪ್ರಯೋಗಾಲಯಕ್ಕೆ ಅಗತ್ಯವಿರುವ ಸಲಕರಣೆ ಮತ್ತು ಬಿಡಿಭಾಗಗಳನ್ನು ಸರಬರಾಜು ಮಾಡುವ ಕುರಿತು" ಎಂದು ನಮೂದಿಸತಕ್ಕದ್ದು.
2. ದರಪಟ್ಟಿಯ ಅವಧಿಯು ದರಪಟ್ಟಿಯನ್ನು ತೆರೆದ ದಿನಾಂಕದಿಂದ 30 ದಿನಗಳ ಅವಧಿಯವರೆಗೆ ಚಾಲ್ತಿಯಲ್ಲಿರಬೇಕು.
3. ಸರಬರಾಜು ಸಂಸ್ಥೆಯವರು ತಮ್ಮ ಸಂಸ್ಥೆಯ GSTIN No. ಮತ್ತು PAN No. ದಾಖಲೆಗಳನ್ನು ದರಪಟ್ಟಿಯ ಜೊತೆಗೆ ಲಗತ್ತಿಸಬೇಕು.
4. ಸರಬರಾಜು ಸಂಸ್ಥೆಯವರು ಅಧಿಕೃತ ಮಾರಾಟಗಾರರು/ಸರಬರಾಜುದಾರರು/ತಯಾರಕರು ಆಗಿದ್ದಲ್ಲಿ ದಾಖಲೆಗಳನ್ನು ಲಗತ್ತಿಸುವುದು.
5. ಸರಬರಾಜು ಮಾಡಿದ 30 ದಿನಗಳೊಳಗೆ ಹುಂಡಿಯನ್ನು ಪಾವತಿಸಲಾಗುವುದು.
6. ದರಪಟ್ಟಿಯನ್ನು ಸಲ್ಲಿಸಲು ಕೊನೆಯ ದಿನಾಂಕ 23.08.2019 ಸಂಜೆ 5-30.
7. ದರಪಟ್ಟಿಯನ್ನು ದಿನಾಂಕ: 26.08.2019 ರಂದು ಮಧ್ಯಾಹ್ನ 12-30ಕ್ಕೆ ತೆರೆಯಲಾಗುವುದು. ದರಪಟ್ಟಿಗಳನ್ನು ಕುಲಸಚಿವರು, ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ಕಾರ್ಯಾಲಯ, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ಬಿ.ಹೆಚ್. ರಸ್ತೆ, ತುಮಕೂರು-572 103 ಇವರಿಗೆ ಮುದ್ದಾಂ ಅಥವಾ ಅಂಚೆಯ ಮೂಲಕ ತಲುಪಿಸಬಹುದು. ಅಂಚೆಯ ವಿಳಂಬಕ್ಕೆ ಅಥವಾ ಅಂಚೆಯು ತಲುಪದಿದ್ದಲ್ಲಿ ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಜವಾಬ್ದಾರಿಯಾಗುವುದಿಲ್ಲ.
8. ದರಪಟ್ಟಿಗಳನ್ನು ಯಾವುದೇ ಕಾರಣ ನೀಡದೆ ಅಂಗೀಕರಿಸುವ ಅಥವಾ ಖರೀದಿ ಪ್ರಕ್ರಿಯೆಯ ಯಾವುದೇ ಹಂತದಲ್ಲಿ ತಿರಸ್ಕರಿಸುವ ಪೂರ್ಣ ಹಕ್ಕನ್ನು ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಕಾಯ್ದಿರಿಸಿಕೊಂಡಿದೆ.
9. ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಒಂದು ವೇಳೆ ತಮ್ಮ ದರಪಟ್ಟಿಯನ್ನು ತಾತ್ಕಾಲಿಕವಾಗಿ ಅಂಗೀಕರಿಸಿದ್ದಲ್ಲಿ ಮಾದರಿ ಸಲಕರಣೆ ಮತ್ತು ಬಿಡಿಭಾಗಗಳನ್ನು ಪ್ರಚುರಪಡಿಸತಕ್ಕದ್ದು.
10. ತಮ್ಮ ದರಪಟ್ಟಿಯಲ್ಲಿ ಪ್ರತಿ ಸಾಮಗ್ರಿವಾರು ದರಗಳನ್ನು ನಮೂದಿಸುವುದು. (ಎಲ್ಲಾ ತೇರಿಗೆ ಮತ್ತು ಸರಬರಾಜು ವೆಚ್ಚ ಒಳಗೊಂಡಿರತಕ್ಕದ್ದು.)
11. ತಮ್ಮ ಸಂಸ್ಥೆಯು ಸರಬರಾಜುಮಾಡಲು ಉದ್ದೇಶಿಸಿರುವ ಸಲಕರಣೆ ಮತ್ತು ಬಿಡಿಭಾಗಗಳ ಸಂರಚನೆಗಳಲ್ಲಿ ವ್ಯತ್ಯಾಸ ಇದ್ದರೆ ಸಂರಚನೆಗಳ ಮಾಹಿತಿ/ಮಾದರಿ ಸಂಖ್ಯೆಗಳನ್ನು ದರಪಟ್ಟಿಯೊಂದಿಗೆ ಕಡ್ಡಾಯವಾಗಿ ಸಲ್ಲಿಸುವುದು.
12. ಈ ದರಪಟ್ಟಿಯು ರೂಪಾಯಿ ಒಂದು ಲಕ್ಷಕ್ಕೆ ಮೀತಿಗೊಳಪಟ್ಟಿರುತ್ತದೆ.
13. ಸರಬರಾಜು ಆದೇಶದಲ್ಲಿ ನೀಡುವ ಎಲ್ಲಾ ಷರತ್ತುಗಳನ್ನು ಕಡ್ಡಾಯವಾಗಿ ಪಾಲಿಸಬೇಕು.
14. ಹೆಚ್ಚಿನ ವಿವರಗಳು ಅಥವಾ ಮಾಹಿತಿಯು ಅಗತ್ಯವಿದ್ದಲ್ಲಿ ಕುಲಸಚಿವರು, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ತುಮಕೂರು ಇವರನ್ನು ಸಂಪರ್ಕಿಸಬಹುದು.

ಪ್ರತಿಗಳು:

1. ಸಿಸ್ಟಮ್ ಅನಾಲಿಸ್ಟ್, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ತುಮಕೂರು. (ಇಂದೇ ವೆಬ್‌ಸೈಟ್‌ನಲ್ಲಿ ಪ್ರಕಟವಾಗುವ ವಿಶ್ವವಿದ್ಯಾನಿಲಯ)
2. ರಕ್ಷಣಾ ಪ್ರತಿ.


ಕುಲಸಚಿವರು
ಕುಲಸಚಿವರು
ತುಮಕೂರು

ಅನುಬಂಧ-01

ಕ್ರ.ಸಂ.	ಸಾಮಗ್ರಿಗಳ ಹೆಸರು	ಸಾಮಗ್ರಿಗಳ ಸಂರಚನೆ	ಪ್ರಮಾಣ	ಕಂಪನಿ ಹೆಸರು ಇದ್ದಲ್ಲಿ ನಮೂದಿಸುವುದು ಅಥವಾ ಇಲ್ಲದಿದ್ದಲ್ಲಿ ಕಂಪನಿ ರಹಿತ ಎಂದು ನಮೂದಿಸುವುದು.	ಪ್ರತಿ ಒಂದಕ್ಕೆ ಗಲನ್ನು ನಮೂದಿಸುವುದು. (Price Per Unit)
1	Digital multimeter	(3 3/4AUTO+TEMP) Display : Large LCD Display with Auto/Manual Range AC Voltage Range : 0.1mV ~ 750V AC Voltage Accuracy : +/- (0.8%+ 6digit) DC Voltage : 0.1mV ~ 1000V DC Voltage Accuracy : +/- (0.5% + 4digit) AC Current Range : 0.1uA ~ 20A AC Current Accuracy : +/- (1.5% + 5digit) DC Current Range : 0.1uA ~ 20A DC Current Accuracy : +/- (1.0% + 5digit) Resistance : 0.1ohm ~ 40Mohm Resistance Accuracy : +/- (0.8% + 2digit) Capacitance : 10pF ~ 200uF Capacitance Accuracy : +/- (3.5% + 8digit) Frequency : 0.1Hz ~ 30MHz Frequency Accuracy : +/- (0.5% + 4digit) Temperature : -40 oc to 1000 oc Temperature Accuracy : +/- (0.8% + 4digit) Count : 4000 Key Features: 3 3/4 Digital Multimeter 4000 Counts Large LCD Display with Auto/Manual Range No Power- FF under natural operation Data Hold, Max. / Min. Value Hold Capacitance, Frequency / Duty Cycle, Temperature and Transistor Test.	10		
2	Lux Meter	Range : 0~2,00,000 LUX Key Features : DATA HOLD Auto Power FF PEAK HOLD(LX-102A) Warranty : 6 Months HTC Lux Meter LX-101A 2 00 000 LUX Model HTC-II Lux Meter LX-101A 2 00 000 LUX	02		
3	Digital clock and timer	Display: 100 mm x 40 mm Specifications: Both up and Down Count Direction Accuracy: 1/100 Sec Large Digits With On/Off Switch In Colored outter plastic box with see-through cover With On/Off Button	05		
4	K-Type Sensor	HITSAN INCORPORATION 100cm Wire Measuring Temperature K-Type Sensor Probe Tester	12		

5	Diffraction Spectrometer	Spectrometer with 6" Circle: Adjustable Prism table marked with grooved lines: Telescope mounted on movable pillar having fine adjustment: Double ending verniers reading 30 Secs Arc: Collimator mounted on fixed pillar with objective	02		
6	Cutting Plier	Insulated Lineman Combination Cutting Plier, Taparia: Red /green color insulator	02		
7	Nose plier	Long Needle Nose Mini Plier Taparia Red /green color insulator	02		
8	Screw Driver Set	Screw Driver Set with Neon Bulb Taparia: Red /green color insulator	03		
9	Grating	Analytical Student Grating Glass Or Diffraction Grating Glass 600 Lines I.E 15000 LPI (Indian)	02		
10	Transistor	NPN, BC547	60		
11	Transistor	Silicon 2N3904	40		
12	Mouse	MS116 Optical Mouse (Black)	04		
13	Magnetic stirrer	Highly Polished Stainless Steel Top Plate for Optimum Heat Transfer Housing: MS Powder Coated :Maximum Heating Temperature up to 200°C. Heating Temperature Range Ambient - 350° Speed Range: 0 to 2200 rpm. :Accurate Step Less Speed Control Maintains Speed Stability :Stirring Capacity up to 2 liters On / Off Power Switch	02		
14	Green board	4X4 ft (Length X Breadth)	02		
15	Thermometer	Mercury Bulb, Range:0-250o C:1/1	10		
16	Logic gates	ICs 7400	10		
17		ICs 7404	10		
18		ICs 7408	10		
19		ICs 7432	10		
20	Op-Amp	IC741	40		
21	Batteries	Multimeter Batteries: 9V out put	50		
22	Temperature Sensor	RTD (Pt-100) Size: 6mm diameter x 100mm length with 2m cable Range: 0-100 deg C Price each: Rs 650/-	03		

23	Constant current source	230 V AC, 50 Hz $\pm$ 10%. Operating Temp. Range: 0°C to 50°C. Current changes by 1% when the load changes till the voltage on the load reaches a value of 24 V Current should read in milliamps on the panel meter. Low current range: 1 mA to 20 mA. Current remains constant to 5 parts in 103 as the load is varied till the voltage across the load reaches a value of 15 V.	01		
24	Digital Temperature Indicator	Mains 230V AC operation, seven segment LED Display (12.5mm height, red) Input: RTD (Pt-100) Range: 0-100 deg C; Resolution: 0.1 deg C	01		
25	Spike busters	Grounds AC power with 3-line protection through all 8 sockets Maximum Spike Current: 13,000 Amps Delivers power through a 2-metre, heavy duty cables	01		


 ಕುಲಸಚಿವರು
 ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾಲಯ
 ತುಮಕೂರು