

ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ

ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ಕಾರ್ಯಾಲಯ, ಬಿ.ಎಚ್.ರಸ್ತೆ, ತುಮಕೂರು-572 103
ದೂರವಾಣಿ : 0816-2254546, 0816-2255596, 0816-2255597 ಫ್ಯಾಕ್ಸ್ : 0816-2270719

ಸಂಖ್ಯೆ:ತು.ವಿ.ಖ.ಕೋ:104:2025-26/ ೨೭೪೬

ದಿನಾಂಕ:08.12.2025

ದರಪಟ್ಟಿ ಆಹ್ವಾನ

- ವಿಷಯ: ವಿಶ್ವವಿದ್ಯಾನಿಲಯದ ಸ್ನಾತಕೋತ್ತರ ಸಾವಯವ ರಸಾಯನಶಾಸ್ತ್ರ ವಿಭಾಗದ ಪ್ರಯೋಗಾಲಯಕ್ಕೆ ಅಗತ್ಯವಿರುವ ಉಪಕರಣಗಳನ್ನು ಸರಬರಾಜು ಮಾಡಲು ದರಪಟ್ಟಿಗಳನ್ನು ಆಹ್ವಾನಿಸಿರುವ ಕುರಿತು.
- ಉಲ್ಲೇಖ: 1. ಸಂಯೋಜಕರು, ಸ್ನಾತಕೋತ್ತರ ರಸಾಯನಶಾಸ್ತ್ರ ಅಧ್ಯಯನ ಮತ್ತು ಸಂಶೋಧನ ವಿಭಾಗ, ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ವಿಜ್ಞಾನ ಕಾಲೇಜು ಇವರ ಪತ್ರ ದಿನಾಂಕ:19.11.2025
2. ವಿ.ವಿ ಅಧಿಸೂಚನೆ ಪತ್ರದ ಸಂಖ್ಯೆ: ತು.ವಿ.ತು.ಕೆ.ಟಿ.ಪಿ.ಪಿ/ತಿ.ಪ/ಖ.ಕೋ/2022-23/9035
ದಿನಾಂಕ: 09.09.2022
3. ಮಾನ್ಯ ಕುಲಪತಿಗಳ ಅನುಮೋದನೆ ದಿನಾಂಕ: 03.12.2025

* * * * *

ಮೇಲ್ಕಂಡ ವಿಷಯಕ್ಕೆ ಸಂಬಂಧಿಸಿದಂತೆ, ಉಲ್ಲೇಖ(1)ರ ಬೇಡಿಕೆಯನುಸಾರ ಉಲ್ಲೇಖ(2)ರ ಆದೇಶದ ಮೇರೆಗೆ ಕರ್ನಾಟಕ ಪಾರದರ್ಶಕತೆ ಅಧಿನಿಯಮ 2000ರ ತಿದ್ದುಪಡಿ 4(i)(ii) ರ ನಿಯಮದನ್ವಯ ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ಸ್ನಾತಕೋತ್ತರ ಸಾವಯವ ರಸಾಯನಶಾಸ್ತ್ರ ವಿಭಾಗದ ಪ್ರಯೋಗಾಲಯಕ್ಕೆ ಅಗತ್ಯವಿರುವ ಪ್ರಯೋಗಾಲಯದ ಉಪಕರಣಗಳನ್ನು ಸರಬರಾಜು ಪಡೆಯಲು ಈ ಕೆಳಕಂಡ ಷರತ್ತು ಮತ್ತು ನಿಬಂಧನೆಗಳಿಗೊಳಪಟ್ಟು ನೋಂದಾಯಿತ ಸಂಸ್ಥೆಗಳಿಂದ ದರಪಟ್ಟಿಗಳನ್ನು ಆಹ್ವಾನಿಸಲಾಗಿದೆ.

List of Equipments with Specifications

Sl. No.	Name of the Equipment	Details/ Specifications	Qty. Req.	Price Per Unit (including tax)
1	Digital Colorimeter Model EQ-651A	Make :Equiptronics or Equivalent One button, μ Controller based colorimeter with 10 inbuilt filters Function : % T & OD Filters: Inbuilt filters, GaN based. 400, 420, 470, 500, 530, 560, 590, 620, 660, & 700 nm. Selection is done using soft touch keys. Readout : 3 digit LED display for %T & OD 3 digit LED display for Wavelength Measurement : %T - 0-100% OD - 0-1.99 Accuracy : %T - $\pm 1\%$ OD - ± 0.01 Light source : LED of infinite life Detector : Photo cell Cuvette : Square APH cuvette 10 mm path length optiglass window Power required : 230 V AC $\pm 10\%$ 50 Hz, 10 VA/ INBUILT RECHARGEABLE BATTERY Warning time : 5 minutes Weight : 1 Kg. (Approx) Body : NANO Crystalline ABSTM Dimension : 90 mm(H) x 225 mm(W) x 220 mm(L) Sample quantity : 1ml Accessories : Square cuvette, Dust proof cover	01	
2	Digital Conductivity Meter Model EQ-664A	MAKE : Equiptronics or Equivalent Auto Ranging Conductivity Meter with inbuilt magnetic stirrer & standard conductance (useful for titration) Selects the right working range Readout : 4 digit LED display Unit of meas. : SIEMENS (MHO) Range : 0.001μS to 200 nS in six auto ranges Six ranges : 200 mS, 20mS, 2mS, 200 μS, 20 μS, 2 μS Accuracy : $\pm 1\% \pm$ last 1 digits Power required : 230 V AC $\pm 10\%$ @50 Hz, 6 VA Weight : 1 Kg. (Approx.) Body : NANO Crystalline ABSTM Warming period : 5 minutes Dimension : 90 mm (H) x 225 mm (W) x 220 mm (L) Standard cond. : Conductance of 1.000 mS Electrode:EQ-708A Unbreakable APH Sleeved Conductivity Cell(K=1) Platinized Platinum Stirrer speed : 500 RPM Fixed Accessories : Teflon magnet, Electrode clamp, Rod, Screwdriver, and Dust proof cover Technology : Quantum Zuerkst Certified	01	

3	pH cum Potentiometer +magnetic stirrer Model EQ-631	MAKE : Equiptronics or Equivalent Readout : 4 digit LED display Range : 0 to 14.00 pH Resolution : pH: 0.01pH Accuracy : pH $\pm 0.01 \pm 1$ least count Power required: 230 V AC $\pm 10\%$ @ 50 Hz, 4 VA. Inbuilt Rechargeable battery Warming period : 5 minutes Weight : 2 Kg Body : ABS Stirrer Speed : 500 RPM fixed Dimension : 100 mm (H) x 230 mm (W) x 240 mm (L) Accessories : Stand set, Electrode Clamps, patch cord For electronic buffer, buffer tablets for 9.2, 7.00, 4.00 pH Screwdriver, Dust proof cover & Telfon Magnet. Electrode : EQ-700 Unbreakable APH Sleeved Glass + Reference Combined Electrode Electronic : Equivalent signal output of ideal electrode Buffer in solution of 9.2, 7.00, 4.00 pH useful for self-diagnostic POTENTIOMETER No need of Calibration Readout : 4 digit LED display Range : ± 1.999 V Accuracy : 0.001 V Electrode : EQ-703 Pt + reference combined (KCL filled)	01	
4	Weighing Balance PBG:321	Make : WENSAR or Equivalent Capacity :330 g Readability:0.001 g (1 mg) Repeatability ($\pm$) :0.001 g Linearity ($\pm$):0.003 g Pan Size :90 mm Response Time:3-5 Sec Display LCD with black light Calibration: Automatic External Tare Range: Full Operating Temperature:10°C to 40°C Power Supply :AC 220V / 50HZ Dimension (LxWxH)mm: 360 x 260 x 355 Weight:3.5.kg	01	
5	Centrifuge Models:R-83	Remi or Equivalent Capacity:400 ml Rotor Capacity(ml):16x15ml Type of Head: Angle rotor head No. on Tubes:16 Weight:30 Kg Digital Timer Range:0-60 minutes Max.Speed:5250 RPM Max. RCF: 3600 g W x D x M : 365x415x350mm Supply:220-240 Volts 50 Hz Single Phase	01	
6	Magnetic Stirrers with hot plate Model:2MLH	Make : Remi or Equivalent SALIENT FEATURES PMDC motor ofr higher torque even at low speeds Better speed regulation even with small volume and low speeds Accurate stepless speed control maintains excellent speed stability. Digital Speed Indicator for displaying of stirring speed Totally enclosed unit Designed for use even in corrosive atmosphere Technical Specifications Stirring Capacity:2 Ltrs Heating Capacity: 300 Watts External Dimensions: 200 x 225 x 185 mm Stirring Paddle (PTFE Coated) - Model:Q-19A, $\varnothing$ x length 9x35mm	02	
7	Sand bath	Sand bath 360x360 mm up to 350°C Variable wattage control Robust construction Eloxidised aluminium heating surface Good uniformity over the entire heating surface, including the corners and edges Optional digital controller 2200watt, 230V Dimensions:360x220x360 mm WxHxD	01	

ಷರತ್ತು ಮತ್ತು ನಿಬಂಧನೆಗಳು:

1. ಮುಚ್ಚಿದ ಲಕೋಟೆಯ ಮೇಲೆ "ವಿಶ್ವವಿದ್ಯಾನಿಲಯದ ಸ್ನಾತಕೋತ್ತರ ರಸಾಯನಶಾಸ್ತ್ರ ವಿಭಾಗದ ಪ್ರಯೋಗಾಲಯಕ್ಕೆ ಅವಶ್ಯವಿರುವ ಉಪಕರಣಗಳ ಸರಬರಾಜಿಗಾಗಿ ದರಪಟ್ಟಿ" ಎಂದು ನಮೂದಿಸತಕ್ಕದ್ದು.
2. ದರಪಟ್ಟಿಯನ್ನು ಸಲ್ಲಿಸಲು ಕೊನೆಯ ದಿನಾಂಕ:20.12.2024 ಅಪರಾಹ್ನ 3.30 ಗಂಟೆ.
3. ದರಪಟ್ಟಿಗಳನ್ನು ಕುಲಸಚಿವರು, ವಿಶ್ವವಿದ್ಯಾನಿಲಯ ಕಾರ್ಯಾಲಯ, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ಬಿ.ಹೆಚ್. ರಸ್ತೆ, ತುಮಕೂರು-572 103 ಇವರಿಗೆ ಮುದ್ದಾಂ ಅಥವಾ ಅಂಚೆಯ ಮೂಲಕ ತಲುಪಿಸಬಹುದು. ಅಂಚೆಯ ವಿಳಂಬಕ್ಕೆ ಅಥವಾ ಅಂಚೆಯು ತಲುಪದಿದ್ದಲ್ಲಿ ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಜವಾಬ್ದಾರರಾಗುವುದಿಲ್ಲ.
4. ತಮ್ಮ ದರಪಟ್ಟಿಯಲ್ಲಿ ಪ್ರತಿ ಸಾಮಗ್ರಿವಾರು ದರಗಳನ್ನು ನಮೂದಿಸುವುದು. (ಎಲ್ಲಾ ತೆರಿಗೆ ಮತ್ತು ಸರಬರಾಜು ವೆಚ್ಚ ಒಳಗೊಂಡಿರತಕ್ಕದ್ದು.)
5. ದರಪಟ್ಟಿ ಸಲ್ಲಿಸುವ ಸಂಸ್ಥೆಯವರು ತಮ್ಮ ಸಂಸ್ಥೆಯ GSTIN No. ಮತ್ತು PAN No. ದಾಖಲೆಗಳನ್ನು ದರಪಟ್ಟಿಯ ಜೊತೆಗೆ ಲಗತ್ತಿಸಬೇಕು.
6. ದರಪಟ್ಟಿ ಸಲ್ಲಿಸುವ ಸಂಸ್ಥೆಯವರು ಅಧಿಕೃತ ಮಾರಾಟಗಾರರು/ಸರಬರಾಜುದಾರರು/ತಯಾರಕರು ಆಗಿದ್ದಲ್ಲಿ ದಾಖಲೆಗಳನ್ನು ಲಗತ್ತಿಸುವುದು.
7. ದರಪಟ್ಟಿಗಳನ್ನು ಯಾವುದೇ ಕಾರಣ ನೀಡದೆ ಅಂಗೀಕರಿಸುವ ಅಥವಾ ತಿರಸ್ಕರಿಸುವ ಪೂರ್ಣ ಹಕ್ಕನ್ನು ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಕಾಯ್ದಿರಿಸಿಕೊಂಡಿರುತ್ತದೆ.
8. ದರಪಟ್ಟಿಯಲ್ಲಿ ನಮೂದಿಸಿರುವ ಉಪಕರಣಗಳ ಪ್ರಮಾಣವನ್ನು ಹೆಚ್ಚಿಸುವ ಅಥವಾ ಕಡಿಮೆಗೊಳಿಸುವ ಹಕ್ಕನ್ನು ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಕಾಯ್ದಿರಿಸಿಕೊಂಡಿರುತ್ತದೆ.
9. ದರಪಟ್ಟಿಯನ್ನು ಅಂಗೀಕರಿಸುವಾಗ ಸಂಸ್ಥೆಗಳು ಸಲ್ಲಿಸಿದ ದರಪಟ್ಟಿಗಳಲ್ಲಿನ ಉಪಕರಣವಾರು ಅತೀ ಕಡಿಮೆ ದರಗಳನ್ನು ವಿಶ್ವವಿದ್ಯಾನಿಲಯವು ಪರಿಗಣಿಸಲಾಗುವುದು.
10. ತಮ್ಮ ಸಂಸ್ಥೆಯು ಸರಬರಾಜು ಮಾಡಲು ಉದ್ದೇಶಿಸಿರುವ ಪ್ರಯೋಗಾಲಯದ ಉಪಕರಣಗಳ ಸಂರಚನೆಗಳಲ್ಲಿ ವ್ಯತ್ಯಾಸ ಇದ್ದರೆ ಸಂರಚನೆಗಳ ಮಾಹಿತಿ/ಮಾದರಿ ಸಂಖ್ಯೆಗಳನ್ನು ದರಪಟ್ಟಿಯೊಂದಿಗೆ ಕಡ್ಡಾಯವಾಗಿ ಸಲ್ಲಿಸುವುದು.
11. ಪ್ರತಿ ಸಂಸ್ಥೆಯವರು ಏಕೈಕ ದರಪಟ್ಟಿಯನ್ನು ಮಾತ್ರ ಸಲ್ಲಿಸತಕ್ಕದ್ದು.
12. ಸರಬರಾಜು ಆದೇಶದಲ್ಲಿ ನೀಡುವ ಎಲ್ಲಾ ಷರತ್ತುಗಳನ್ನು ಪಾಲಿಸಬೇಕು.
13. ಹೆಚ್ಚಿನ ವಿವರಗಳು ಅಥವಾ ಮಾಹಿತಿಯು ಅಗತ್ಯವಿದ್ದಲ್ಲಿ ಉಪಕುಲಸಚಿವರು, ಖರೀದಿ ಕೋಶ, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ತುಮಕೂರು ಇವರನ್ನು ಸಂಪರ್ಕಿಸಬಹುದು.

ಆದೇಶದ ಮೇರೆಗೆ


ಉಪಕುಲಸಚಿವರು
ಉಪಕುಲಸಚಿವರು
ಖರೀದಿ ಕೋಶ

ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ

ತುಮಕೂರು

ಪ್ರತಿಗಳು:

1. ಸಿಸ್ಟಮ್ ಅನಾಲಿಸ್ಟ್, ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ತುಮಕೂರು. (ಇಂದೇ ವೆಬ್‌ಸೈಟ್‌ನಲ್ಲಿ ಪ್ರಕಟಿಸತಕ್ಕದ್ದು).
2. ಅಧ್ಯಕ್ಷರು, ಸ್ನಾತಕೋತ್ತರ ಸಾವಯವ ರಸಾಯನಶಾಸ್ತ್ರ ಅಧ್ಯಯನ ಮತ್ತು ಸಂಶೋಧನಾ ವಿಭಾಗ ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ, ಜ್ಞಾನಸಿರಿ ಕ್ಯಾಂಪಸ್, ಬಿದರಕಟ್ಟೆ, ತುಮಕೂರು ತಾಲ್ಲೂಕ್.
3. ರಕ್ಷಣಾ ಪ್ರತಿ.