

No.TUT/PC/2025-26/

Dated:24.03.2026

**Corrigendum**

The Tender Notification was issued through KPPP Portal vide this office **TU/PC/PMUSHA/EQ/2025-26** Dated: **05.03.2026** the same was published in the News Paper, "Indian Express" and "Prajapragati" Tender Title supply and Installation of Scientific Equipments and others Sanctioned Under PM-USHA Grants. As per pre-bid meeting held on 18.03.2026, herewith a corrigendum is issued for the bidders. Bidders are requested to read the following amendments:

Item No.	Item Name	Specification As in Tender Document	Read as
PMUE_001	FT-IR Spectrophotometer with ATR and UPS for analysis of organic samples	<b>FTIR Specifications with Diamond ATR:</b>	Computer with software support is required to run the Instrument
PMUE_002	Digital Balance (4 digit)	Capacity – 252 g	Weighing Capacity 220 g
PMUE_003	Analytical Balance	Readability 0.0001 g (1 mg)	Readability 0.001
PMUE_029	Hot Air Oven	Size 18" x 18" x 18" Inches	Capacity 220 Ltr do not consider the size
PMUE_032	Soxhlet extractor	Material – Mild Steel, Temperature Range – Up to 300 °C, Phase – Single Phase, Display Type – Digital, Power Source – Electric	No. of position 01 Flask capacity 500 ml without glassware
PMUE_037	Digital weighing Balance (10 mg to 300 g)	#Capacity 300 g-Minimum display 0.001 g #Repeatability (Std. Deviation) ≤0.1 mg #Linearity ±0.2 mg #Stabilisation Time *1 Approx. 3.0 sec. #Operating Temperature and Humidity Limits: 5-40°C 20-85% #Temperature Coefficient for Sensitivity (10-30°C)±2ppm/°C #Power requirement: 12V, 1A	Minimum display 0.01 g
PMUE_041	Lab deep freeze	LAB DEEP FREEZER (-20° C) HORIZONTAL (MODEL NO. HV-DF-121) by hoverlabs	Volume capacity 200 Ltr
PMUE_058	TLC Plates plastic based silica gel	Advance instruments make	TLC Chamber for plate size 10 x 10 cm
PMUE_063	Centrifuge machine	Hettich MIKRO 200/200R Centrifuge Parameter:15000rpm,21382rpm AC/DC input: SCHUKO Plug,200-240V AC,Euro Plug,UK plug H X W X D 260mm(10.25 in, X 275mm (10.75in,)X 344mm(13.5in,)	Non-refrigerated
PMUE_072	Semi Automatic Microphone	Section Thickness Indication : Visual LCD display. Range of section thickness : 0.5 - 99 µm in 1 µm steps. Coarse specimen advance : 1 - 99 µm(adjustable). Horizontal specimen feed : 25mm. Vertical specimen feed : 60 mm. Specimen Orientation : XY axis : +/-8 Degrees / Z axis : 360 Degrees. Convenient specimen Holder : Standard 55 x 55 mm.	Semi Automatic Microtome

(P.T.O)

PMUE_075	Remi Magnetic stirrer-5Lts Hotplate (5-ML)	Capacity 500ml, 2 Litres Brand-Remi Operating Frequency-50 Hz Power-150 - 500 W Stirring capacity-5 Ltrs Stirring paddle model-Q - 20 A	Without Hotplate
----------	--	---	------------------

**Note:** Other Conditions, Specifications remains unchanged.

Tumkur University, Tumkur reserves the right to accept/reject/cancel any or all tender schedules without assigning any reason.

  
Deputy Registrar  
Tumkur University  
Tumkur

ಉಪಕುಲಸಚಿವರು  
ಐರೀದಿ ಕೋಶ  
ತುಮಕೂರು ವಿಶ್ವವಿದ್ಯಾನಿಲಯ  
ತುಮಕೂರು

**Copy to:**

The System Analyst, Tumkur University, Tumkur. – Publish in the E-procurement Portal.